

CERTIFICATE OF CALIBRATION

for the Nasal Ranger[®] Field Olfactometer

Serial Number : 90201690

Calibration Date : 7/10/2014

| Dial D/T | Actual D/T | % Variance |
|-----------------|-------------------|-------------------|
| 60 | 59.81 | -0.3% |
| 30 | 30.03 | 0.1% |
| 15 | 15.07 | 0.5% |
| 7 | 7.00 | 0.0% |
| 4 | 4.00 | 0.0% |
| 2 | 2.00 | 0.0% |

This document certifies this Nasal Ranger[®] Field Olfactometer, specified by unique Serial Number, was calibrated using a NIST traceable primary gas flow standard by St. Croix Sensory, Inc.

**St. Croix Sensory, Inc.
1150 Stillwater Blvd. N.
Stillwater, MN 55082 USA
+1-651-439-0177
info@nasalranger.com**



Michelle Arty
Calibration Technician